


LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO.,LTD.
NO.1111, SOUTH HUIHE ROAD, CHAOYANG DISTRICT, BEIJING, CHINA 100020

The following samples were submitted and identified by/on behalf of the client as:

Name of Product/Item : NITRILE EXAMINATION GLOVES
Country of Origin : CHINA
Sample Receiving Date : JUN.21,2018
Testing Period : JUN.21,2018 TO JUL.09,2018
Test Performed : SELECTED TEST(S) AS REQUESTED BY APPLICANT
Test Requested : EN 455-1:2000 MEDICAL GLOVES FOR SINGLE USE –
PART 1: REQUIREMENTS AND TESTING FOR FREEDOM FROM
HOLES
Test Result(s) : FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING
PAGE(S)
Conclusion : THE SUBMITTED SAMPLE MET THE TEST REQUIREMENT.

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Approved by:



Zhou Xinkuan, SK
Lab Manager

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

QDHL 001259



SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Hardlines

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
中国·青岛·崂山区株洲路143号通标中心 邮编: 266101

t (86-532) 68999187 f (86-532) 80991952 www.sgs.com.cn
t (86-532) 68999187 f (86-532) 80991952 e sgs.china@sgs.com

Test Conducted:

EN 455-1:2000 Medical gloves for single use – part 1: Requirements and testing for freedom from holes

Number of test sample	:	200 Pieces
The type of gloves	:	examination/procedure gloves
Manufacturing batch code	:	/
Batch size	:	/
Sample size	:	M
Number of non-conforming gloves	:	0
Defects observed before testing	:	No defects
Test Result	:	Pass

Clause	Test Items	Result	Note
5	Watertightness test for detection of holes	---	---
5.1	Referee testing	Pass	# 1

Notes : #1 See result 1

Test Result:

1. Watertightness test for detection of holes

Sample Quantity: 200

AQL: 1.5 Accept: 7 Reject: 8 Found: 0

Remark: The sample selecting amount for Watertightness test for detection of holes is deviated to 200 pcs as accessed by SGS.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Sample Photo:

Received sample



SGS authenticate the photo on original report only

End of Report

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

QDHL

001261